



California Medical Device Recall Information



Recall Name

Animas Recalls the 2020 Insulin Infusion Pump Due to a Possible Malfunction

Recall Date	Product Description	Recalling Firm	Recall Reason
1/03/12	Animas 2020 Insulin Pump	Animas Corp. West Chester, PA	<i>Potential for pumps to sound a false alarm or warning; and a software limitation that prevents the pump from delivering insulin past 12/31/2015</i>
Recall Class	Product Identification	Distribution	Affected Dates
I	Animas 2020 Insulin Pump <ul style="list-style-type: none">Suspected recalled product include all lots manufactured from March 1, 2012 to November 30, 2012	CA , nationwide	Distributed from March 1, 2012 to September 30, 2012

FOR ADDITIONAL INFORMATION, PLEASE VISIT:

<http://www.fda.gov/MedicalDevices/Safety/ListofRecalls/ucm346650.htm>